

PACKAGE 33: SUPPLY AND DELIVERY OF MEDICAL EQUIPMENT & MATERIALS

Item #	Item/Description	UOM	Total Qty	Unit Cost (ABC)	Total Amount
1	<p>HOSPITAL BED, PLAIN WITH MATTRESS</p> <ul style="list-style-type: none"> - Size: 2030 x 900 x 500mm - epoxy coated steel framework - stainless steel head and foot bows; foot bow with DOH logo printed in the center, color: green measuring 1" x 1/2" - mounted on protective stumps - provision for IV rod (IV socket) ; 4 (one on each side) - with side rails (at least 1/2 length) aluminum alloy - with patient name holder at foot and head part - knock-down construction - 4" mattress, leatherette cover with DOH letters printed in the center (front), color green measuring 6" x 1" <p>Standard requirements:</p> <ul style="list-style-type: none"> - current and valid certificate of manufacturer's compliance with ISO 13485 - user's manual in English Language <ul style="list-style-type: none"> a. operations manual 2 copies b. service manual 2 copies - the brand has been in the local market for at least ten (10) years) <p>After Sales Service:</p> <ul style="list-style-type: none"> - bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five(5) years after the warranty period <p>Warranty Certificate:</p> <ul style="list-style-type: none"> - two (2) year warranty for parts and services after the acceptance by the user. Warranty period must start on the date of acceptance by the end-user 	unit	64	30,000.00	1,920,000.00

PACKAGE 33: SUPPLY AND DELIVERY OF MEDICAL EQUIPMENT & MATERIALS

Item #	Item/Description	UOM	Total Qty	Unit Cost (ABC)	Total Amount
2	<p>PATIENT LITTER, BRANDED AND BRAND NEW</p> <p>Standard Features: Technical Specifications: - minimum dimensions: width: 550+/-20mm length: 1950mm (+20mm, -50mm) height: 750mm - loading capacity: at least 120 kgs - the stretcher must have a quick release patient restrain system -frame material: high grade aluminum - frame must be made of sturdy light weight, non-twisting construction and with rubber handles - stretcher bed material must be either cloth or canvass with DOH print on the center mobility, portability; handled 2 pcs restrain strap with carrying case</p> <p>Standard requirements: - current and valid certificate of manufacturer's compliance with ISO 13485 - user's manual in English Language a. operations manual 2 copies b. service manual 2 copies - the brand has been in the local market for at least ten (10) years)</p> <p>After Sales Service: - bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five(5) years after the warranty period Warranty Certificate: - two (2) year warranty for parts and services after the acceptance by the user. Warranty period must start o the date of acceptance by the end-user</p>	unit	1002	3,000.00	3,006,000.00

PACKAGE 33: SUPPLY AND DELIVERY OF MEDICAL EQUIPMENT & MATERIALS

Item #	Item/Description	UOM	Total Qty	Unit Cost (ABC)	Total Amount
3	<p>WEIGHING SCALE W/ HEIGHT MEASURING STICK, ADULT, BRANDED AND BRAND NEW</p> <p>Technical Specifications:</p> <ol style="list-style-type: none"> 1. Mechanical weighing scale with height rod 2. 180 kgs max. weighing capacity 3. Non-Slip steel platform, at least 10" width x 14" length 4. Adjustable/ telescopic height rod with at least 5mm gradation 5. Height rod can measure height from 30" to at least 75" 6. Engraved numbers on eye level scale beam and height rod 7. Scale reads in both metric (kg.G) and imperial (lbs) units 8. Weight increments/ graduation of 50 grams 9. Height rod or aluminum material with plastic head positioned 10. Powder coated body 11. Sharp edge construction 12. Built-in transport castor <p>Standard Requirements:</p> <ol style="list-style-type: none"> 1. Users manual in English Language <ol style="list-style-type: none"> a. Operations Manuals 2 copies b. Service Manual 2 copies 2. Valid certificate of manufacturer's compliance with ISO 13485 3. The brand has been in the local market for at least ten (10) years 4. Copy of distributorship agreement between the manufacturer and the local distributor (bidder) 5. List of completed projects in the last five (5) years with copies of contract or purchase orders attached for verification purposes; 6. Marketing authorization, registration approval of free sale certificate of the product issued by the health authority in the country of origin <p>After Sales Service:</p> <ol style="list-style-type: none"> 7. Certification that the bidder will provide a Service Unit that the end-user can use in case the equipment will be pulled-out for repair or maintenance within the warranty period 8. List of engineers or technical personnel who shall perform preventive maintenance of the product during the warranty period <p>Warranty Service:</p> <ol style="list-style-type: none"> 9. Two (2) years warranty for parts and services after the acceptance by the end-user; 10. Location, contact number of the service center of the supplier (bidder) 	unit	1002	10,000.00	10,020,000.00

PACKAGE 33: SUPPLY AND DELIVERY OF MEDICAL EQUIPMENT & MATERIALS

Item #	Item/Description	UOM	Total Qty	Unit Cost (ABC)	Total Amount
	11. Certification that the bidder/ supplier has the engineer or technician trained to conduct the corrective and preventive maintenance of the product. The service engineer or technician must be presently employed by the bidder/ supplier or authorized by the manufacturer 12. Certification that the manufacturer shall guarantee the availability of spare in the next five (5) years;and 13. Metrological certification of calibration from the manufacturer				

TOTAL 14,946,000.00