

PACKAGE 34 - SUPPLY AND DELIVERY OF OTHER EQUIPMENT

| Item # | Item/Description | Unit | Total Qty | Unit Cost (ABC) | Total Amount |
|--------|--|------|-----------|-----------------|--------------|
| 1 | <p>Autoclave, 200 L, brand new, branded</p> <p>Technical Specifications:</p> <ul style="list-style-type: none"> • With full installation and commissioning • Automatic operation • Steam Floor Type Autoclave Sterilizer; Sterilization Time 0-60 min; Horizontal Type Sterilizer, Round Chamber; Approx. 200-Liter Chamber Capacity • Operating Temperature: 118-134 degrees Celsius adjustable ; Material: stainless Chamber and Jacket; Auto-Dry Function; Temperature Display; Auto-fill Water System; Digital Temperature Controller; Pressure Safety Valve ; Chamber Pressure Display ; Over pressure protection; Jacket Pressure Display; Water level control switch; Program (Function) Display ; Emergency Exhaust Switch; Indicator Light for Power, Add-Water, Heat; Sterilization, Dry, Complete; Program Selection for Wrapped, Unwrapped, Liquid; • Power Supply 220 V –single phase, 60 Hz <p>Standard Accessories</p> <ul style="list-style-type: none"> • 1 piece Stainless steel tray; 2 pcs stainless tong; Autoclave test pack (10); With trolley and sliding tray <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Current and Valid Certificate of Manufacturer’s compliance with ISO 13485 and/or CE Certificate • Installation and commissioning; • Original Brochure; User’s manual in English Language; Operations Manual 2 copies in English Language at least 1 original colored copy; Service Manual 2 copies in English Language <p>Training requirements</p> <ul style="list-style-type: none"> • Training of at least two (2) end-users; <p>After Sales Service</p> <ul style="list-style-type: none"> • Bidder’s certificate that parts shall be available for a period of five (5) years after the warranty period. <p>Warranty Certificate</p> <ul style="list-style-type: none"> • One (1) year for parts and service after inspection, testing and acceptance • Delivery: on site of the receiving facility within sixty (60) calendar days upon receipt of the Notice to Proceed (NTP). | unit | 2 | 520,000.00 | 1,040,000.00 |

PACKAGE 34 - SUPPLY AND DELIVERY OF OTHER EQUIPMENT

| Item # | Item/Description | Unit | Total Qty | Unit Cost (ABC) | Total Amount |
|--------|--|------|-----------|-----------------|--------------|
| 2 | <p>Power Generating Set 6 KVA Capacity: 10 KVA; Voltage: 220 Volts Phase: Single Phase Frequency: 60 Hz General control panel: Display Analog or Digital Panel control with Voltmeter, Battery Voltmeter, Electric Start Engine: Gasoline Fuel tank: Fuel tank base mounted tank with a capacity of at least 2 hours operation full. General Requirements: Material and workmanship – All materials and parts of the electric generator set shall be brand new. Standard Requirements: • Current and Valid Certificate of Manufacturer’s compliance with ISO standards or its equivalent Standard Requirements: • User’s Manual in English Language; (a) Operations Manual 2 copies, at least 1 original colored copy; (b) Service Manual 2 copies Training requirements • Training of at least two (2) end-users After Sales Service: •Bidder’s certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period. Warranty Certificate: •One (1) year for parts and two (2) years on the required service after delivery, inspection, testing, acceptance and after training of end user has been completed. • Delivery: on site of the receiving facility within sixty (60) calendar days upon receipt of the Notice to Proceed (NTP).</p> | Unit | 1 | 100,000.00 | 100,000.00 |

PACKAGE 34 - SUPPLY AND DELIVERY OF OTHER EQUIPMENT

| Item # | Item/Description | Unit | Total Qty | Unit Cost (ABC) | Total Amount |
|--------|--|------|-----------|-----------------|--------------|
| 3 | Tracheostomy Set, brand new, branded Standard Features: Made of non-toxic PVC, transparent, soft and smooth. • Permanently attached 15mm connector. Pre-connected necktape. • Size (mm): 5.0, 5.5, 6.0, 6.5, 7.0, 7.5, 8.0, 8.5, 9.0, 9.5, 10.0 Standard Requirements: • Current and Valid Certificate of Manufacturer’s compliance with ISO 13485 and/or CE Certificate • User’s Manual in English Language 2 copies Warranty Certificate • Minimum of one (1) year Warranty Certificate for craftsmanship, upon delivery, inspection and acceptance • Delivery: on site of the receiving facility within sixty (60) calendar days upon receipt of the Notice to Proceed (NTP) | UNIT | 10 | 5,000.00 | 50,000.00 |

TOTAL 1,190,000.00